



Agilent E1420 VXI*plug&play* Instrument Driver Read Me

README.TXT for HP E1420 Counter

Contents:

1 Introduction

2 VTL/VISA Specific Information

2.1 Instrument Addresses

2.2 Executing the Soft Front Panel without VTL/VISA

3 Using the HP E1420 Driver in Application Development Environments

3.1 C/C++

3.2 Microsoft Visual Basic

4.0 Known Defects

5.0 Online Information

6.0 Revision History

1. Introduction

The help file included with the HP E1420 VXI*plug&play* driver contains instrument and programming information. This readme file contains additional information for the HP VXI*plug&play* programmer.

2. VTL/VISA Specific Information

The following topics refer to instances when the driver is used with different versions of VTL.

2.1 Instrument Addresses

When using HP VXI*plug&play* instrument drivers, instrument addresses must

be all uppercase letters. Implementation of the addressing scheme is vendor specific and some vendors support mixed cases. However, for maximum portability, the instrument address should use uppercase characters only.

For example: use "VXI0::56::INSTR" instead of "vxi0::56::instr"

2.2 Executing the Soft Front Panel without VTL/VISA

The VTL/VISA dynamic link library provides a communication library between your program and the instrument. The SFP may be executed without VISA or without the HPE1420.dll present. It will do a partial emulation of the hardware for demo purposes.

3. Using the HP E1420 Driver in Application Development Environments

This section offers suggestions on using the hpe1420.dll within various application development environments.

3.1 C/C++

Refer to your C/C++ manuals for information on linking and calling .DLLs.

Rebuilding the driver DLL should be done in a different directory than the one the driver was installed in order to differentiate the changes.

3.2 Microsoft Visual Basic

Refer to the Microsoft Visual BASIC manual for information on calling DLLs. The BASIC include file is hpe1420.bas, which is contained in the directory ~vxipnpwinXXinclude. The ~ refers to the directory in the VXIPNP variable. By default this is set to C:. winXX refers to the operating system -- win95 or winNT.

4.0 Known Defects

5.0 Online Information

The latest copy of this VXIplug&play driver and other HP

VXIplug&play and SCPI drivers are available at the Hewlett-Packard customer WWW site: http://www.hp.com/go/inst_drivers

If you do not have Internet access, the drivers can be obtained by contacting your country support organization. In the US, the phone number is 800-452-4844.

A list of contact information for non-US countries is available on the Access HP website: <http://www.tmo.hp.com/tmo/contacts/English/callcenters.html>

6.0 Revision History

A.01.00 May 1, 1996 Initial Release.

A.02.00 Aug 1996

A.02.01 2/98

hpe1420_32.dll: A.02.01

hps1420.exe : A.02.01

A.02.02 11/98

hpe1420_32.dll: A.02.02

hps1420.exe : A.02.02

A.02.03 8/99

hpe1420_32.dll: A.02.02

hps1420.exe : A.02.02

PRINT

CLOSE