



Certificate of Calibration

TEST RESULTS:

The resistor (UUT) was calibrated by using a Binary Voltage Divider Resistance Bridge, Model 6000, to compare it with a calibrated 100,000,000 Ω resistor in an air bath maintained at 23.00°C +/- 100mk.

The UUT was allowed to stabilize a minimum of 24 hours in a temperature controlled bath. Each Voltage run had a settling rate of 45 seconds, 10 measurements with 9 measurements for statistics. . This was repeated 6 times, the type A uncertainty for the standard deviation of each measurement and the spread of values being inserted into the Uncertainty Analysis worksheet. The type B uncertainty for the measurement comes from the uncertainty of the 100M Ω resistor. The type B and type A uncertainties are root sum squared and doubled to give expanded uncertainty.

The reported value of resistance is based on the results found when the UUT was in circulating air.

| Resistance (MOhms) | Uncertainty ($\mu\Omega/\Omega$) | Voltage (V) | Bath Temperature (°C) |
|-----------------------|---------------------------------------|----------------|--------------------------|
| 100.00877 | 8.5 | 20 | 23 |

MEAN DATE OF MEASUREMENT: June 25, 2022



Certificate of Calibration

CERTIFICATE NO.:

C1220622

CUSTOMER NAME:

xDevs.com

CUSTOMER ADDRESS:



MEASURAND:

MODEL NO.: SRL-100M

S/N.: K1-1309646

MFG.: IET

DESCRIPTION: 100 Mega Ohm Standard Air Resistor

CALIBRATION RANGE(S) OR POINTS COVERED BY THIS CERTIFICATE:

The measurement was performed with a test voltage of 20V at a bath temperature of 23C.

CALIBRATION PROCEDURE

CAL-11-011-04

REFERENCE STANDARD(S):

| MFG. | DESCRIPTION | MODEL NO. | S/N. | CALIBRATED DATE | CERTIFICATE NO. |
|----------------------------|------------------------------------|------------|---------|-----------------|-----------------------------|
| Measurements International | 100 Mega Ohm Standard Air Resistor | 9331G/100M | 1100179 | April 28, 2021 | 2021010047-1 (RM 34.276) |

ENVIRONMENTAL CONDITIONS:

AMBIENT:
TEMPERATURE: 23 °C ± 2 °C
HUMIDITY: 41 % ± 10 %
BAROMETRIC PRESSURE: 101 kPa

OF MEASURAND:
TEMPERATURE: 23.00 °C ± 0.05 °C
HUMIDITY: 41 % ± 10 %

UNCERTAINTY OF MEASUREMENT

THE UNCERTAINTY OF MEASUREMENT IS ESTIMATED TO BE:

THE REPORTED UNCERTAINTY OF MEASUREMENT IS STATED AS THE COMBINED STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR OF $k = 2$. THE MEASURED VALUE (y) AND THE ASSOCIATED UNCERTAINTY (U) REPRESENT THE INTERVAL ($y \pm U$) WHICH CONTAINS THE MEASURED QUANTITY WITH A PROBABILITY OF APPROXIMATELY 95%. THE UNCERTAINTY WAS ESTIMATED USING ISO GUIDE TO THE EXPRESSION OF UNCERTAINTY IN MEASUREMENT (GUM) GUIDELINES. THE ESTIMATED UNCERTAINTY CONTAINS CONTRIBUTIONS ORIGINATING FROM THE MEASUREMENT STANDARD CALIBRATED BY A NATIONAL LABORATORY, FROM THE CALIBRATION METHOD, FROM THE ENVIRONMENTAL CONDITIONS AND FROM THE MEASURAND BEING CALIBRATED. THE LONG-TERM BEHAVIOUR OF THE MEASURAND IS NOT INCLUDED.

AMENDMENTS (IF APPLICABLE):

CERTIFICATE NO. AMENDED: _____ REASON FOR AMENDMENT: _____

CALIBRATED BY (SIGNATURE)

DATE OF CALIBRATION

June 24 - 25, 2022

AUTHORIZING SIGNATURE

DATE OF ISSUE

22/06/27

The reported measurements contained within this report relate only to the measurands calibrated.
These measurements are traceable to national standards and thus to the International System of Units (SI).